

AMENDMENT NO. 1 JULY 1975
TO
IS:4244-1967 SPECIFICATION FOR
RETRACTORS, SKIN, DOUBLE ENDED

Alterations

(Page 5, clauses 6.1 and 6.1.1) - Substitute the following for the existing clauses:

'6.1 Corrosion Resistance - The instrument shall satisfy the boiling and autoclaving test as mentioned in IS:7531-1975*.'

(Page 5, foot-note) - Insert the following foot-note at the bottom of the page:

Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments

(CPDC 11)

Reprography Unit, ISI, New Delhi