

AMENDMENT NO. 1 JULY 1975

TO

IS:4245-1967 SPECIFICATION FOR NEEOLE
HOLDER (GILLIES) COMBINED WITH
SCISSORS, FOR PLASTIC SURGERY

Alterations

(Page 7, clauses 6.4 and 6.4.1) - Substitute
the following for the existing clauses:

'6.4 Corrosion, Resistance_ - The instrument shall
satisfy the boiling and autoclaving test as
mentioned in IS:7531-1975*.'

(Page 7, foot-note) - Insert the following foot-
note at the bottom of the page:

'*Method for boiling and autoclaving test for
corrosion resistance of stainless steel surgical
instruments.'

(CPDC 11)

Reprography Unit, ISI, New Delhi.