

AMENDMENT NO. 1 MARCH 1971
TO
IS : 4380-1967 SPECIFICATION FOR ABNEY LEVEL

Alterations

(*Page 4, clause 3.2*) — Substitute the following for the existing clause:

3.2 The vernier shall be truly concentric and there shall be no gap between the main scale and the vernier. '

(*Page 5, clause 4, heading*) —

Substitute ' **TECHNICAL REQUIREMENTS** '
for ' **GENERAL REQUIREMENTS** '.

(*Page 5, clause 5.1, line 3*) — Substitute ' If the error exceeds 10 minutes ' for ' If the difference is appreciable '.

(*Page 6, clauses 6.1 and 6.2*) — Substitute the following for the existing clauses:

' **6.1** The weight of the instrument shall not exceed 250 g.

6.2 The weight of the carrying case of the instrument shall not exceed 150 g. '

Addendum

(*Page 6, clause 5.3.4*) — Add the following new clause after 5.3.4 and re-number the existing clauses '5.4 and 5.5' as '5.5 and 5.6' respectively:

' **5.4** The reflectivity of the reflector shall not be less than 65 percent. '