

AMENDMENT NO. 1 OCTOBER 1980
TO
IS : 4381-1967 SPECIFICATION FOR
PATHOLOGICAL MICROSCOPE

Alterations

(*Page 6, clause 6.1.3*) — Substitute the following for the existing clause:

'**6.1.3** In changing from lower magnification to next higher magnification or re-introducing the same objective by rotation of the nosepiece. the object at the centre of the field shall be well within the field of view.'

(*Page 6, clause 6.1.6*) — Substitute the following for the existing clause:

'6.1.6 Parfocality of the objectives and the eyepieces shall be within half a turn of the line motion knob except for 6X and lower power objectives and eyepieces '

(*Page 7, clause 6.2.3*) — Substitute the following for the existing clause:

'**6.2.3** When observed under white light illumination, the image shall appear clear and well defined The image shall be totally free from colour defects up to half the held of view and shall be reasonably free from colour up to two-thirds held of view '

(*Page 7, clause 6.2.4, first sentence*) — Substitute the following for the existing sentence.

'The image should be reasonably free from curvature of field to the extent of two-thirds of the total field.'