

**AMENDMENT NO. 2 FEBRUARY 2004
TO
IS 4381 : 1967 SPECIFICATION FOR
PATHOLOGICAL MICROSCOPE**

(*Page 4, clause 4.5*) —Substitute the following for the existing Note 1:

' Note 1 The value of numerical aperture stipulated above are the nominal.'

(*Page 8, clause 6.2.8, last sentence*) — Substitute the following for the existing sentence:

'The measured values shall not differ from the specified values by ± 5 percent.'

(CHD 35)

Reprography Unit, BIS, New Delhi, India