

AMENDMENT NO. 1 JULY 1975
TO
IS:4490-1968 SPECIFICATION FOR FORCEPS,
DILATING, TROUSSEAU'S

Alterations

(Page 5, clauses 6.2 and 6.2.1) - Substitute the following for the existing clauses:

'6.2 Corrosion Resistance - The instrument shall satisfy the boiling and autoclaving test as mentioned in IS : 7531-1975*.'

(Page 5, foot-note) - Insert the following foot-note at the bottom of the page:

*Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments.'

(CPDC 11)

Reprography Unit, ISI, New Delhi