

AMENDMENT NO. 2 JULY 2021

TO

IS 16112 : 2013 BETA PICOLINE — SPECIFICATION

(Page 2, Annex A, A-2.3) — Substitute ‘Column — general purpose non polar capillary GC column containing 100 percent dimethyl polysiloxane (PDMS) or capillary GC column containing high polarity polyethylene glycol (PEG) phase’ for ‘Column — CP-Sil-52CB Wax, 60 m long, 0.25 mm internal diameter, 0.25 µm film thickness (df).’