

TO
IS:5194-1969 METHOD FOR DETERMINATION
OF NITROGEN-KJELDAHL METHOD

Addendum

(Page 4, clause 4.2) - Add the following
note after 4.2:

'Note - In order to avoid back suction of the
liquid in the beaker, presence of positive pressure
by introduction of gas (nitrogen gas or air free
from carbon dioxide) would make the operation
smoother.'

(CDC 1)

Reprography Unit, BIS, New Delhi, India