

AMENDMENT NO. 1    DECEMBER 1979

TO

IS:5204-1969    SPECIFICATION FOR RESEARCH MICROSCOPE

Alterations

(Page 5, clause 4.6) - Substitute the following for the existing clause:

'4.6 Light Source - The microscope shall be equipped with a high intensity compact light source enclosed in a suitable housing complete with condensing lens, iris diaphragm and filters including ground glass for critical and Köhler illumination of the specimen.'

(Page 7, clause 4.9.1.1, line 2) - Substitute '0.01 mm' for '0.001 mm'.

(EDC 36)

---

Reprography Unit, ISI, New Delhi