

AMENDMENT NO. 2 SEPTEMBER 1980
TO
IS : 5204-1969 SPECIFICATION FOR
RESEARCH MICROSCOPE

Alterations

[*Page 6, clause 4.8.1(a) to (d)*] — Substitute the following for the existing matter:

- a) 32 mm ($5 \times \text{NA } 0.125$),
- b) 16 mm ($10 \times \text{NA } 0.25$),
- c) 4 mm ($40 \times \text{NA } 0.65$) or ($45 \times \text{NA } 0.65$), and
- d) 2 mm ($90 \times \text{NA } 1.25$) or ($100 \times \text{NA } 1.25$) oil immersion.'

(*Page 9, clause 6.1.3*) — Substitute the following for the existing clause:

'6.1.3 In changing from lower magnification to next higher magnification or re-introducing the same objective by rotation of the nosepiece, the object at the centre of the field shall be well within the field of view.'

(*Page 10, clause 6.2.6*) — Substitute the following for the existing clause:

'6.2.6 When observed under white light illumination, the image shall appear clear and well defined. The image shall be totally free from colour defects up to half the field of view and shall be reasonably free from colour up to two-thirds field of view.'

Addenda

(*Page 6, clause 4.8.2*) — Add the following new note after 4.8.2:

'NOTE — The mechanical tube length of 160 mm is applicable only in case of monocular microscope while for bi-nocular microscope the tube length may be up to 200 mm.'

(*Page 9, clause 6.1.1*) — Add the following new note after 6.1.1:

'NOTE — The mechanical tube length of 160 mm is applicable only in case of monocular microscope while for bi-nocular microscope the tube length may be up to 200 mm.'