



AMENDMENT NO. 3 MAY 1983

TO

IS:5204-1969 SPECIFICATION FOR RESEARCH MICROSCOPE

Alterations

(Page 10, clause 6.2.4):

- a) *heading* - Substitute the following for the existing heading:

*'Field of View Number (Field Number) of Eyepiece'.*

- b) *Last sentence* - Substitute the following matter for the existing:

*'An alternative method for the measurement of field of view number shall be as given below:*

A microscope fitted with an objective with magnification 40X or 45X and the eyepiece under test shall be focused on a stage micrometer. The exact magnification of the objective shall be determined according to the method given in Appendix J of IS:2754-1964\*. A stage micrometer normally has 1 mm divided into 100 divisions, thereby making each graduation equal to 0.01 mm. The maximum number of graduations on the stage micrometer which are seen with the eyepiece under test are noted. This number expressed in mm gives the diameter of the visual field of the microscope. The field of view number of the eyepiece is given by the formula:

|               |                               |
|---------------|-------------------------------|
|               | Diameter of the visual        |
| Field of view | = field in mm X magnification |
| number        | of the objective.'            |

(EDC 36)

---

Reprography Unit, ISI, New Delhi, India