

AMENDMENT NO. 2 AUGUST 1992
TO
IS 5291 : 1969 SPECIFICATION FOR TABLES,
OPERATION, HYDRAULIC, MAJOR

(Page 4, clause 3.4) — Insert the following at the end of the clause:

‘or else Designation **04Cr18Ni10** or **07Cr17Ni12Mo2** in accordance with IS 6911: 1972’.

(Page 4, clause 3.5) — Substitute the following for the existing clause and the Note:

‘35 Antistatic rubber tread castors **according** to IS 4034 : 1979 or cotton impregnated, bakelite wheels shall be used. The construction of the table shall however ensure transmission of the static charge to the earth.’

(Page 4, clause 4.1, last sentence) — Substitute the following for the existing **sentence**:

‘The overall width of the table top excluding side railings shall be 50 to 55 cm.’

(Page 9, clause 8.1) — Substitute the following for the existing clause:

‘8.1 The table shall be provided with the following accessories:

- a) Foot extension — 1 No.
- b) Shoulder supports — 2 No.
- c) Lateral supports or side supports — 2 No.
- d) Anaesthetic **screen** — 1 No.
- e) Kidney elevator — 1 No.
- f) Arm supports — 1 No.
- g) Pelvic holder strap — 1 No.
- h) **Leg** strap — 2 No.
- j) Wristlets — 2 No.
- k) Pairs of knee **crutches/Lithotomy** leg holders — 2 No.
- m) Intravenous arm board — 1 No.
- n) 25 to **50** mm thick antistatic foam rubber mattress — 1 set.