

AMENDMENT NO. 1 JUNE 2006
TO
IS 1223 : 2001 APPARATUS FOR DETERMINATION
OF MILK FAT BY GERBER METHOD —
SPECIFICATION

(Third Revision)

(Page 14, Section 6, clause 19.2.2) — Insert the following at the end of the existing text:

'In case of a hand operated centrifuge, the temperature of the contents of the butyrometer shall be taken after centrifuging by inserting a thermometer in a butyrometer after taking the reading of the fat column.'

(Page 14, Section 6, clause 19.4.3, line 1) — Substitute 'Electric driver centrifuge' for 'Centrifuge'.

(Page 14, Section 6, clause 19.4.3, line 7) — Insert the following at the end of the existing text:

'In case of a hand operated centrifuge, speed is controlled by hand only and appropriate speed indicator shall be provided for governing the same.'

(FAD 19)