

AMENDMENT NO. 1 SEPTEMBER 2025
TO
IS 9471 (PART 6) : 2024 MODULAR LOWER LIMB ORTHOTIC COMPONENTS —
SPECIFICATION
PART 6 ORTHOTIC HIP JOINT (JOINT UNIT, HIP DROP LOCK)
(First Revision)

(Page 1, clause 5.2) — Insert the following after **5.2**:

‘5.3 Joint hip lower shall be supplied in Zn plated to service condition 1 as per IS 1573.’

[Page 2, Table under Fig. 1, Sl No. (iv)] — Substitute the following for the existing:

<i>‘ Sl No.</i>	<i>Dimension</i>	<i>Size I</i>	<i>Size II</i>	<i>Size III</i>
(1)	(2)	(3)	(4)	(5)
iv)	DIA, D $\pm$ 0.25	7	9	9 ,