

AMENDMENT NO. 1 APRIL 1999
TO
IS 9489 : 1980 METHOD OF TEST FOR THERMAL
CONDUCTIVITY OF THERMAL INSULATION
MATERIALS BY MEANS OF HEAT FLOW METER

(*Page 8, clause 3.8*) — Substitute the following for the existing text:

‘A potentiometer or any other suitable equipment having a sensitivity of 0.1 percent of minimum output from the thermocouple or better and an accuracy of ± 0.1 percent or better shall be used for measurement of all thermocouples and thermopile emf.’

(CHD 27)

Reprography Unit, BIS, New Delhi, India