

AMENDMENT NO. 2 JULY 2009
TO
IS 14439 (PART 2) : 1998 LEGAL METROLOGY —
GAS VOLUME METERS

(Page 3, *clause 8.2.1.2*) — Substitute the following for the existing clause:

‘At flowrates equal to or greater than $0.1 Q_{\max}$ the errors shall be determined independently at least six times, by varying the flowrate between each consecutive measurement. The difference between any two errors found at each test flowrate shall not exceed 0.6 ‘percent’