

**AMENDMENT NO. 2 JANUARY 1996
TO
IS 7083 : 1973 SPECIFICATION FOR TROLLEY,
MEDICINE**

(Page 1, clause 3.3) — Add the following matter at the end of the clause:

“or mild steel tube of 15 mm diameter conforming to IS 1239 (Part 1) : 1990
‘Mild steel tubes, tubulars and other wrought steel fittings: Part 1 Mild steel
tubes (*fifth revision*)’.”

(MHD 14)

Printed at : Prabhat Offset Press, New Delhi-2