

AMENDMENT NO. 1 SEPTEMBER 1980
TO
IS : 7115-1973 SPECIFICATION FOR APPARATUS, TUBAL PATENCY
Alteration

(*Page 2, Clause 5.1*) — Substitute the following for the existing clause:

" **5.1 Corrosion Resistance Test (Applicable Only to Stainless Steel Cannula)** — Test the cannula in accordance with IS : 7531-1975 'Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments'. The cannula shall show no sign of corrosion after the test. "

(CPDC 25)