

AMENDMENT NO. 1 NOVEMBER 1935

TO

IS : 7324-1983 SPECIFICATION FOR BRX HYDROMETERS

*(First Revision)*

*(Page 5, clause 7)* - Add the following new clause after 7:

"7.0 The brix hydrometers and brix thermo-hydrometers shall conform with the design requirements given in IS:9621-1980 'Principles of construction and adjustment of glass hydrometers'."

*(Page 7, Fig. 1A)* - Substitute '335 Max' for '335 Min'.

*(Page 8, Table 2, Sl No. 1, col 2)* - Substitute 'Max' for 'Min'.

(CDC 33)