

AMENDMENT NO. 1 JANUARY 2024
TO
IS 3674 : 2020 TEXTILES — COTTON FIBRES — DETERMINATION OF
MICRONAIRE VALUE
(First Revision)

(Page 3, Table 1) — Substitute the following for the existing:

Table 1 Quality Characteristics of ICAR-CIRCOT Calibration Cottons
(Clause A-1.1)

SI No.	Code	Micronaire Value
(1)	(2)	(3)
i)	Low mic	2.6 to 2.9
ii)	High mic	5.0 to 5.6
iii)	Tolerance	± 0.1
NOTES		
1 The label of each calibration cotton shows the exact test values of each parameter and lot number. The values of the different parameters of each calibration cotton differ based on the lot number.		
2 The tolerance shall be applicable on the exact test values of the lot as indicated on the label of calibration cotton.		