

**AMENDMENT NO. 1      JULY 1975**  
**TO**  
**IS:3767-1966      SPECIFICATION FOR HOOK,**  
**SKIN, GILLIES      PATTERN**

**Alterations**

*(Page 5, clauses 7.3 and 7.3.1)* - Sub-  
stitute the following for the existing  
clauses:

'7.3 Corrosion Resistance - The instrument  
shall satisfy the boiling and autoclaving  
test as mentioned in IS:7531-1975\*.'

*(Page 5, foot-note)* - Insert the  
following foot-note at the bottom of the  
Page:

'\*Method for boiling- and autoclaving  
test for corrosion resistance of stainless  
steel surgical instruments.'

(CPDC 11)

---

Reprography Unit, ISI, New Delhi