

AMENDMENT NO. 2 AUGUST 1998
TO
IS 3829 (PART 3) : 1985 SPECIFICATION
FOR STEAM STERILIZERS
PART 3 PRESSURE STERILIZERS, VERTICAL
CYLINDRICAL TYPE

(*Page 5, clause 5.1, informal table, col 2, first para*) — Substitute the following with the existing:

'Stainless steel conforming to designation X04Cr19Ni9 or X02Cr19Ni10 and of 2B finish in accordance with IS 6911:1992'.

[*Page 8, clause 9.1(b)*] — Substitute '60 cm deep' *for* '50 cm deep'.

[*Page 8, clause 9.1(c)*] — Substitute '50 cm dia' *for* '40 cm dia'.

(MHD 14)