

**AMENDMENT NO. 2 DECEMBER 1992
TO
IS 3831 : 1979 SPECIFICATION FOR STERILIZER,
SHALLOW (DRESSING DRUM)**

(Second Revision)

[Page 1, clause 2 (i), col 3] — Substitute '2D' for '2B'.

[Page 1, clause 2 (ii), col 4] — Substitute ' 1.00 ± 0.1 ' for '1.00'.

[Page 1, clause 2 (iv)] — Delete.

[Page 1, clause 2 (vi), col 4] — Substitute '3 nominal' for '2.80 to 3.55'.

[Page 1, clause 2 (vi)] — Insert the following new clauses after 2 (vi):

'2.1 Flattening of the hinge wire ends shall be done in such a manner so that hinges do not come out with constant use.'

2.2 A tolerance of ± 0.2 mm on the nominal diameter of the handle shall be permitted.

2.3 A suitable chain minimum 0.9 mm diameter shall be used.'

(Page 1, clause 4.9.1) — Substitute ' 2.4 ± 0.2 ' for '2.4'.

(Page 3, clause 5) — Substitute the following for the existing clause:

'Finish of the different parts shall be as agreed to between the purchaser and the supplier'.

(Page 3, clause 7) — Substitute 'indication of the source of manufacturer' for 'manufacturer's name'.

(Page 3, clause 7.1) — Delete.

(MHD 14)

Reprography Unit, BIS, New Delhi, India