

**AMENDMENT NO. 4 JANUARY 1996
TO
IS 3831 : 1979 SPECIFICATION FOR STERILIZER,
SHALLOW (DRESSING DRUM)**

(Second Revision)

[Page 2, Informal table, line 5, column 2 (see also Amendment No. 3)]:

a) Substitute dimension '125' for dimension '132'.

b) Add the following note below the table:

'NOTE — Overall dimensions of diameter and height are given for the purpose that the drum fits into the sterilizer.'

[Informal Table, column 3 (see also Amendment No. 3)] — Substitute 'E ± 5' for 'E' in the first row.

(MHD 14)

Reprography Unit, BIS, New Delhi, India