

AMENDMENT NO. 1 MAY 2026
TO
IS 3025 (Part 40) : 2024 METHODS OF SAMPLING AND TEST (PHYSICAL AND
CHEMICAL) FOR WATER AND WASTEWATER
PART 40 CALCIUM
(Second Revision)

[Foreword, para 4, Sl No. (b)] — Substitute ‘spectrometry’ for ‘spectroscopy’.

[Page 1, clause 1, Sl No. (d)] — Substitute ‘spectrometry’ for ‘spectroscopy’.

(Page 2, clause 5.4.4.2) — Substitute ‘5.4.4.3 Patton and reeder, indicator solution’ for ‘5.4.4.2 Patton and reeder, indicator solution’.

(Page 3, clause 5.6) — Substitute the following for the existing:

‘5.6 Calculation

$$\text{mg Ca/l} = \frac{A \times M \times 40.08 \times 1\,000}{V}$$

where

A = volume, in ml, of EDTA solution used for titration;

M = molarity of EDTA used for titration; and

V = volume, in ml, of the sample taken for the test.

$$\text{Calcium hardness (as CaCO}_3\text{), mg/l} = \frac{A \times M \times 100 \times 1\,000}{V}$$

where

A = volume, in ml, of EDTA solution used for titration;

M = molarity of EDTA used for titration; and

V = volume, in ml, of the sample taken for the test.’

(Page 7, clause 8) — Substitute the following for the existing:

‘8 INDUCTIVELY COUPLED PLASMA OPTICAL EMISSION SPECTROMETRY AND
INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY

Calcium can also be determined by inductively coupled plasma optical emission spectrometry with reference to procedure given in IS 3025 (Part 2). Likewise, inductively coupled plasma mass spectrometry with reference to procedure given in IS 3025 (Part 65) can also be used for the determination of calcium in water and wastewater.’

(CHD 36)

Price Group 1

Publication, BIS, New Delhi