

AMENDMENT No. 1 MAY 2026
TO
IS 3025 (Part 45) : 2024 METHODS OF SAMPLING AND TEST (PHYSICAL AND
CHEMICAL) FOR WATER AND WASTEWATER
PART 45 SODIUM AND POTASSIUM
(Second Revision)

[Foreword, para 5, Sl No. (b)] — Substitute ‘spectrometry’ for ‘spectroscopy’.

[Page 1, clause 1, Sl No. (e)] — Substitute ‘spectrometry’ for ‘spectroscopy’.

(Page 1, clause 4) — Substitute the following for the existing:

‘4 SAMPLING AND STORAGE

The sampling and storage shall be done as prescribed in IS 17614 (Part 1) and IS 17614 (Part 3).’

(Page 8, clause 10) — Substitute the following for the existing:

‘10 INDUCTIVELY COUPLED PLASMA OPTICAL EMISSION SPECTROMETRY AND
INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY

Sodium and potassium can also be determined by inductively coupled plasma optical emission spectrometry with reference to procedure given in IS 3025 (Part 2). Likewise, inductively coupled plasma mass spectrometry with reference to procedure given in IS 3025 (Part 65) can also be used for the determination of sodium and potassium in water and wastewater.’