

AMENDMENT NO. 1 MAY 2026
TO
IS 3025 (PART 46) : 2023 METHODS OF SAMPLING AND TEST (PHYSICAL AND
CHEMICAL) FOR WATER AND WASTEWATER
PART 46 MAGNESIUM
(Second Revision)

[Foreword, para 4, Sl No. d)] — Substitute ‘spectrometry’ for ‘spectroscopy’.

[Page 1, clause 1, Sl No. c)] — Substitute ‘spectrometry’ for ‘spectroscopy’.

(Page 2, clause 5.4.7, line 2) — Substitute ‘disodium ethylenediaminetetraacetate dihydrate (EDTA)’ for ‘disodium ethylenediaminetetraacetate dehydrate (EDTA)’

(Page 4, clause 5.6) — Substitute ‘6.6 Calculation’ for ‘5.6 Calculation’.

(Page 4, clause 6) — Substitute the following for the existing:

**‘7 INDUCTIVELY COUPLED PLASMA OPTICAL EMISSION SPECTROMETRY AND
INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY**

Magnesium can also be determined by inductively coupled plasma optical emission spectrometry with reference to procedure given in IS 3025 (Part 2). Likewise, inductively coupled plasma mass spectrometry with reference to procedure given in IS 3025 (Part 65) can also be used for the determination of magnesium.’

(CHD 36)

Price Group 1
Publication, BIS, New Delhi