

**AMENDMENT NO. 1 MARCH 2002
TO
IS 3055 (Part 1) : 1994 CLINICAL THERMOMETERS
PART 1 SOLID STEM TYPE — SPECIFICATION**

(*Second Revision*)

(*Second cover page, Foreword, para 3*) — Insert the following new para after para 3:

"The responsibility of 'first official verification' as mentioned under **8.2.2** and to mark 'a recognized mark of the verifying authority' as mentioned under **9.1** (e) shall be of officers authorized under the provision of the '*Standards of Weights and Measures (Enforcement) Act, 1985*'."

(*Page 1, clause 5.1*) — Insert the following new clause after **5.1**:

'5.2 In addition, the thermometers may also be graduated in degrees Fahrenheit (°F).'

(*Page 3, clause 9.1*) — Insert the following new clause after **9.1** and renumber the subsequent clause:

'9.1.1 Clinical thermometers with dual scale (i.e. Celcius and Fahrenheit) shall be legibly and indelibly marked on its stem with the letters °C and °F near the top of the relevant scale.'

(MHD 12)