

AMENDMENT NO. 2 APRIL 2002
TO
IS 3055 (Part 1): 1994 CLINICAL THERMOMETERS
PART 1 SOLID STEM TYPE — SPECIFICATION

(Second Revision)

(Page 3, clause **8.2.2**) — Substitute the following for the existing clause:

'8.2.2 Clinical thermometers after ageing as per **8.2.1** shall ensure that their accuracy as measured by the method specified in Annex C, shall be as given below for at least one year after the first official verification:

<i>Range °C</i>	<i>Accuracy °C</i>
Less than 35.8	± 0.3
Greater than 35.8 but less than 37.0	± 0.2
37.0 to 39.0	± 0.1
Greater than 39.0 to 41.0	± 0.2
Greater than 41.0	± 0.3'

(MHD 12)