

**AMENDMENT NO. 3 OCTOBER 2004
TO
IS 3055 (PART 1): 1994 CLINICAL THERMOMETERS
PART 1 SOLID STEM TYPE — SPECIFICATION**


(*Second Revision*)

(*Page 1, clause 2.1*) — Delete reference of IS 4825 : 1968.

(*Page 1, clause 7.2.1*) — Insert the following at the end of the existing text:
"except for 'stability of bulb tube' and 'limit of alkalinity'."

(*Page 1, clause 7.2.3*) — Substitute '263.5 μ g' for '263.5 g'.

(*Page 2, clause 7.3.1.1*) — Substitute the following figure for the existing Fig. 1:



(*Page 2, Table 1*):

a) For overall length at SI No. (iii), substitute '100 to 120' for '100 to 115'.

b) For external diameter of stem at SI No. (v) substitute '4 to 7' for '4 to 6'.

(*Page 2, clause 7.5.3*) — Delete 'or 0.15 mm, whichever is less' from the existing text.

(*Page 3, clause 7.5.6*) — Substitute the following for the existing:

'7.5.6 The numbers shall be placed suitably as shown in Fig. 1.'

(*Page 5, Annex C, clause C-1.2*) — Substitute the following for the existing text:

'Calibrated reference thermometer shall have range of 35° to 43°C and least count 0.05°C.'

(HD 35)

Reprography Unit, BIS, New Delhi, India