

AMENDMENT NO. 1 FEBRUARY 1994
TO
IS 3099 (Part 1 and 2) : 1992 MICROSCOPES - SLIPS
AND SLIDES - SPECIFICATION
(Page 1, clause 2.1, line 3) - Substitute ' $n_e 1.5255 \pm 0.0015$ ' for
' $n_e 1.5225 \pm 0.0015$ '.

(LMD 20)

Reprography Unit, BIS, New Delhi, India