

AMENDMENT NO. 4 AUGUST 2008
TO
IS 3099 (PARTS 1 AND 2): 1992 MICROSCOPE SLIPS
AND SLIDES - SPECIFICATION
PART 1 MICROSCOPE SLIPS
PART 2 MICROSCOPE SLIDES
(First Revision)

(Page I, clause 3.1) - Substitute following for the existing:

'3.1 Thickness

Cover slips shall conform to the following thickness ranges:

NO.1 (General purpose) : $0.17 \begin{smallmatrix} 0 \\ -0.04 \end{smallmatrix}$ mm

No. I-H (High performance) : $0.17 \begin{smallmatrix} 0 \\ -0.02 \end{smallmatrix}$ mm

NOTES

I Microscope manufacturers, for purposes of optical design, use 0.17 mm as the combined thickness of cover slips and mounting medium, measured from the top surface of the cover slips to the top surface of the specimen being observed.

2 In addition to the above No. 1 and No. 1-H cover slips, other thicknesses are available, such as $1\frac{1}{2}$ (0.17 $\begin{smallmatrix} +0.02 \\ -0.01 \end{smallmatrix}$ mm) and No.2 (0.17 $\begin{smallmatrix} +0.08 \\ -0 \end{smallmatrix}$ mm) which may be used for some purposes.

Highest optical quality, particularly with large aperture objectives, may not be obtained with these thicknesses (Note 2 is only of a cautionary nature, but does not form part of the standard).'

(Page 3, clause 3.1) - Substitute following for the existing:

'3.1 Thickness

The thickness for microscope slides shall be $1.1 \begin{smallmatrix} +01 \\ -02 \end{smallmatrix}$

NOTE -Thicker and thinner slides are available. Highest optical quality. particularly with large aperture condensers, may not be obtained with these thicknesses.'

(PG 22)