

AMENDMENT NO. 1 FEBRUARY 2003
TO
IS 3104 (PART 1) : 1982 SPECIFICATION FOR
DENSITY HYDROMETERS

PART 1 REQUIREMENTS

(First Revision)

(Page 5, clause **4.2.1**, line 3) — Substitute '0.50 g/ml' for '0.60 g/ml'.

(CHD 10)

Reprography Unit, BIS, New Delhi, India