

AMENDMENT NO. 2 JULY 1994
TO
IS 5594 : 1983 SPECIFICATION FOR WRIST UNIT,
ROTARY, FOR HAND PROSTHESIS
(First Revision)

(Page 2, Table 1, col 2, 3, 4 and 5) — Incorporate a tolerance of ± 2 mm for all the entries under 'Outer Dia' and 'Width'.

[Page 2, clause 4.1(b)] — Delete this sub-clause and renumber the subsequent sub-clauses.

(MHD 10)

Reprography Unit, BIS, New Delhi, India