

AMENDMENT NO. 1 JUNE 1979
TO
IS:5601-1970 SPECIFICATION FOR GOUGES, STILLE
PATTERN, ORTHOPAEDIC

Alteration

(*Page 2, clause 4.4*) — Substitute the following for the existing clause:

'**4.4** The gouge shall be held in one hand firmly with the edge making an angle of 45° on to a long, hard and dry sheep bone of 10 mm thickness approximately. The bone shall be shaped by chiselling, using the hammer conforming to IS : 5585-1970' Specification for matter rubber-capped'. After 20 strokes the head of the gouge shall not become mushroom shaped and the edge shall not show any deformation.'

Addendum

(*Page 1, clause 3*) — Add the following new clause after 3???

"3.1 A deviation of ± 2.5 percent shall be allowed on all dimensions.'

(CPDC 24)