

AMENDMENT NO. 1 MAY 1993
TO
IS 5631 : 1970 SPECIFICATION FOR TROLLEY,
INSTRUMENT, PLAIN AND CURVED

(Page 1, clause 2.2) — Insert the following at the end of the clause:

'or stainless steel sheet of minimum thickness 8 mm conforming to IS 5522 : 1992 '

(Page 1, clause 3, Fig.1 and 2) :

- a) Substitute dia of the castor wheel '50 or 80' for '80'
- b) Delete the word 'glass' wherever appears.
- c) Substitute 'MS/A1 Cap' for 'MS/A1 Disc'.

(Page 1, clause 4.1) — Substitute the following for the existing clause;

'4.1 Framework — The trolleys shall have a welded framework of continuous lengths of mild steel tubes as shown in Fig.1 and 2. The upright members shall be connected together by other set of tubes as shown in the respective figures. These tubes shall be connected to the upright members in such a way so as to form supporting frames for two shelves, made of glass or stainless steel with shaped corners. If glass shelves are used, for avoiding rattling, suitable size rubber studs or buttons shall be employed and the glass shelves pressed on them over the supporting frames. In case of curved instrument trolley the shelves may consist of two equal segments. The upper ends of the upright members shall be sealed with mild steel or aluminum caps. Rest of the details shall conform to the requirements laid down in 4.1 of IS 4033 : 1968.'

(Page 1, clause 4.2, line 1) — Substitute '50 or 80' for '80'.

(Page 1, clause 5) — Delete the word 'glass'.

(Page 1, clause 6.2.1) — Delete the word 'glass'