

AMENDMENT NO. 1 MAY 1994
TO
IS 5901:1970 SPECIFICATION FOR
THERMOGRAPH BIMETALLIC

(*Page 3, clause 3.1, line 2*) — Add 'aluminium alloy' between the words 'metal' and 'or'.

(*Page 3, clause 3.1.1, line 2*) — Substitute 'aluminium alloy' for 'aluminium'.

(*Page 3, clause 3.2*) — Substitute the following for the existing:

'The material used for the remaining components of the instrument shall be non-magnetic, non-corrosive and non-rusting material like stainless steel, brass or similar material with suitable anticorrosive paint. The material used shall be such that it is capable of being finished to the specified dimensions and is not affected by exposure to widely varying climatic conditions both at inland and coastal stations.'

(*Page 6, clause 5.4.4*) — Add the following at the end of the clause:

'Suitable fibre tipped disposable pen can also be used.'

(*Page 6, clause 5.4.5*) — Substitute 'IS 5924 : 1988 (first revision)' for 'IS 5924 : 1970'.