

AMENDMENT NO. 3 MAY 1994
TO
IS 8336 : 1977 SPECIFICATION FOR
THERMOELECTRIC PYRANOMETER

[Page 2, clause **0.2** (see also Amendment No. 1)] — Substitute the following for the existing:

'0.2 Measurements of the total amount of energy received from the sun and sky are of fundamental importance in meteorology and in studies of the exchange and transformation of energy in the earth-atmosphere system and in biology, medicine, agriculture, architecture and industry. The solar radiation which reaches the earth's surface is contained in the range of wavelengths 0.3 to 4 μm (300 - 4 000 nm). The quantity of irradiance falling on unit area of the surface in unit time is expressed in Wm^{-2} . The solar constant, the solar irradiance received by a unit surface and placed normal to the irradiation just outside the earth's atmosphere at mean sun-earth distance has a value of 1369 Wm^{-2} .

[Page 7, clause **5.1.5** (see also Amendment No. 1)] — Substitute ' $\pm 3.0 \text{ Wm}^{-2}$ ' for ' $\pm 0.3 \text{ mW/cm}^2$ ' and ' $5 \text{ to } 9 \mu\text{VW}^{-1}\text{m}^{-1}$ ' for ' $8 \text{ mv/cal/cm}^2/\text{min}$ '.

(Pages 7 and 8, clauses **5.1.7, 5.1.8, 5.1.9, 5.1.10, 5.1.11, 5.2.4 and 5.2.5**)
— Substitute 'per cent' for 'percent'.

[Page 11, clause **7.1.2(b)** (see also Amendment No. 1)] — Substitute ' $\mu\text{vw}^{-1}\text{m}^{-2}$ ' for ' $\text{mV/cal/cm}^2/\text{min}$ '.