

**AMENDMENT NO. 4 AUGUST 2002
TO
IS 8336 : 1977 SPECIFICATION FOR
THERMOELECTRIC PYRANOMETER**

(*Page 10, clause 5.3.4*) — Insert the following new clause after **5.3.4**:

"5.4 Albedo is the ratio of reflected solar radiation to global solar radiation. Two pyranometers have to be used separately for the measurement of these two components of radiation.

For measurement of reflected radiation, a pyranometer has to be installed in such a way that the sensor faces the surface for which albedo has to be measured but does not affect or interfere with incoming solar radiations fall on the measured surface. The body of Pyranometer should be covered properly with some metallic hood painted white. It will protect the instrument from direct heating due to radiation falling on the body. This is essential because the massive metallic body of the pyranometer works as a sink for the thermo-junctions of the thermopile (sensor) which are in good heat contact with the body.

The Pyranometer with hood has to be installed about 1 to 2 m above the surface on an inverted 'L' shaped stand."

(BP 21)

Reprography Unit, BIS, New Delhi, India