



AMENDMENT NO. 2 OCTOBER 1984

TO

IS:8462-1977 SPECIFICATION FOR STERILIZER, PORTABLE,
VERTICAL, PRESSURE TYPE

Alteration

(Page 1, clause 4.1) - Substitute the following
for the existing clause:

'4.1 The sterilizer chamber shall have a circular
cross section with a diameter of 350 ± 5 mm and overall
height between 300 and 325 mm.'

(CPDC 14)

Reprography Unit, ISI, New Delhi, India