

AMENDMENT NO. 1 JUNE 1982
TO
IS:4033-1968 GENERAL REQUIREMENTS FOR
HOSPITAL FURNITURE

Alterations

(Page 4, clause 2.1, line 2) - Substitute
'IS:2039-1981*' for 'IS:2039-1964*'.
'

[Page 4, clause 2.2(b), (c) and (d)] - Substitute
the following for the existing matter:

'b) Stainless Steel Sheet conforming to designation
04Cr 18Ni 10 or 07Cr 18Ni 9 of IS:6911-1972[†].

c) Aluminium sheet conforming to designation
19000, 31000, 31500, 40800, 51000-A, 51000-B,
51300, or 52000 of IS:737-1974[‡].

d) Galvanized steel sheet conforming to IS:277-1977[§].'

(Page 4, clause 4.2, line 2) - Substitute
'IS : 4034 - 1979^Π, for 'IS:4034-1968[¶].'

(Page 4, foot-notes) - Substitute the following
for the existing foot-notes:

'*Specification for steel tubes for bicycle and
allied purposes,*(second revision)*.

[†]Specification for stainless steel plate, sheet
and strip.

[‡]Specification for Wrought aluminium and aluminium
alloys, sheet and strip (for general engineering
purposes) *(second revision)*.

[§]Specification for galvanised steel sheets (plain
and corrugated) *(third revision)*.

^ΠSpecification for Castors for hospital use
(first revision).'

(Page 5, clause 5.2, line 2) - Substitute
'IS:5-1978' for 'IS:5-1961*'.
'

(Page 5, foot-note with '*' mark) - Substitute
the following for the existing foot-note:

'*Specification for colours for ready mixed paints
and enamels,(third revision).'

(CPDC 14)