

AMENDMENT NO. 1 APRIL 1976
TO
IS : 3320-1973 SPECIFICATION FOR
SURGICAL SCALPELS
(*First Revision*)

Alterations

(Page 3, clause 4.4) — Substitute the following for the existing clause:

'4.4 The tip of the blade shall be well defined, central and sharp. There shall be no waviness, jags, feathers, nicks, or other defects on the cutting edge when examined visually.'

(Page 3, clause 6.1) — Substitute the following for the existing clause:

'6.1 The cutting edge of the scalpel blade shall be examined for nicks under a magnification of 10 ×. The edge shall show no nick of size 0.10 mm or more or two nicks of sizes between 0.05 and 0.10 mm. The indentation below 0.05 mm shall not be considered for this test. '

(CPDC 11)