

**AMENDMENT NO. 1 AUGUST 1992
TO
IS 7399:1987 SPECIFICATION FOR KNIFE,
STERNUM, LEBSCHKE'S PATTERN**

(First Revision)

(Page 1, clauses 2 and 2.1) - Substitute the following for the existing

Clauses:

"2 MATERIAL

2.1 The sternum knife with integral handle shall be made from stainless steel of Designation 30Cr13 of IS 6603 : 1972 Specification for stainless steel bars and Flats'.

2.2 Where the handle of sternum knife is not made integral with the working part the material shall be as follows :

- a) *Working part* - Stainless steel conforming to Designation 40Cr13 of IS 6603 : 1972
- b) Handle Stainless steel conforming to Designation 07Cr18Ni9 of IS 6603 : 1972,"

(MHD 16)

???