

AMENDMENT NO. 1 MAY 2025

TO

**IS 7454 : 2024 REHABILITATION EQUIPMENT — WHEELCHAIRS, FOLDING,
ADULT SIZE — SPECIFICATION**

(Second Revision)

(Page 2, Table 1, Sl No. xiii) — Substitute the following for the existing:

'xiii)	Wheel diameter	534 (24"), <i>Min</i> '.
--------	----------------	--------------------------

(Page 3, clause 6.6) — Substitute the following for the existing:

‘6.6 The solid tyres used in castors shall be made of non-marking resilient rubber or ethylene vinyl acetate (EVA) or polyurethane (PU) or thermoplastic polyurethane (TPU) with a minimum hardness of 60 Shore ‘A’.’

(Page 3, clause 6.7.1) — Substitute the following for the existing:

‘6.7.1 The rear wheels shall be of minimum 534 (24") size and shall have rims, tyres, tubes, nipples, spokes and washers conforming to the relevant Indian Standard. Alternatively, the solid tyres used in rear wheel shall be made of non-marking resilient rubber or ethylene vinyl acetate (EVA) or polyurethane (PU) or thermoplastic polyurethane (TPU) with a minimum hardness of 60 Shore ‘A’.

NOTE — In Type 1 wheelchairs, the rear wheels could be as small as the front wheels.