

**AMENDMENT NO. 1 DECEMBER 1980**  
**TO**  
**IS : 7480-1974 SPECIFICATION FOR PERFORATOR, SIMPSON'S PATTERN**  
**Alteration**

( *Page 2, clause 6.1* ) — Substitute the following for the existing clause:

" **6.1 Corrosion Resistance Test** — Test the perforator in accordance with IS : 7531-1975 \* Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments'. The perforator shall show no sign of corrosion after the test."

(CPDC 25)