

AMENDMENT NO. 1 OCTOBER 1984

TO

IS:7625-1975 SPECIFICATION FOR GAG, MOUTH,
DOYEN'S PATTERN

Alteration

(Page 2, clause 6.2) - Substitute the following
for the existing clause:

"6.2 *Corroeion Resistance Test* - The instimment shall
be tested in accordance with the test specified in
IS:7531-1975 'Method for boiling and autoclaving
test for corrosion resistance of stainless steel
surgical instruments'.

It shall not show any sign of corrosion after
the test."

(CPDC 13)

Reprography Unit, BIS, New Delhi, India