

AMENDMENT NO. 2 SEPTEMBER 2025
TO
IS 8086 : 2024 REHABILITATION EQUIPMENT — WHEELCHAIRS, FOLDING,
JUNIOR SIZE — SPECIFICATION
(Second Revision)

(First cover page, Title) — Substitute the following for the existing Hindi title:

‘दिव्यांगजनों के लिए उपकरण — व्हीलचेयर, फोल्डिंग, छोटा साईज़ — विशिष्टि
(दूसरा पुनरीक्षण)’

(Page 2, clause 7.1.1) — Substitute the following for the existing:

‘7.1.1 The backrest shall be firmly secured to the vertical side members and shall be removable. The seat shall be firmly secured to each side of the frame.’

(Page 2, clause 7.3.2) — Substitute the following for the existing:

‘7.3.2 The footrest shall be capable of swinging/rotating manually about its own axis to clear the way when patient enters or leaves the chair. In the raised position, the footrest shall be at an angle of 90° to 120° to its normal horizontal position.’

(Page 2, clause 7.3.3) — Substitute the following for the existing:

‘7.3.3 The footrest swinging/rotation about its axis shall be controlled by a suitable mechanism to remain at the desired position.’

(Page 2, clause 7.3.4) — Deleted

(Page 4, clause 8.2) — Substitute the following for the existing:

‘8.2 All exposed metallic parts including rear wheels shall be stove enamelled (after primary coat) or powder coated or coated using cathodic electrodeposition or plated as agreed to between the purchaser and the supplier. The resulting finish shall be hard and shall not readily chip or flake.’

(MHD 09)

Price Group 1
Publication, BIS, New Delhi