

AMENDMENT NO. 1 OCTOBER 1974
TO
IS : 6720 - 1972 REFERENCE TABLES FOR PLATINUM/
30 PERCENT RHODIUM - PLATINUM/6 PERCENT
RHODIUM THERMOCOUPLES

Addendum

(Page 3, clause 2.2) — Add the following new clause after 2.2:

'3. TOLERANCES

3.1 The calibration of the thermocouple shall agree with the values given in Table 1 and Table 2 within the following tolerances:

<i>Temperature Range, °C</i>	<i>Tolerance</i>
0 to 600	± 3°C
Above 600	± 0.5 percent '

(ETDC 48)