

AMENDMENT NO. 1 MAY 1994
TO
IS 6805 : 1973 SPECIFICATION FOR ASSMANN
PSYCHROMETER

(*Page 3, clause ????*) — Insert 'Antimagnetic and' between the words 'are' and 'not'.

(*Page 5, clause 5.5*) — Add the following at the end of the clause:

'Suitable electrically driven mechanism can also be used to provided adequate aspiration to the thermometers.'

(*Page 5, clause 6.1, line 3*) — Substitute 'IS 5924 : 1988 (*first revision*)' for '5924 : 1970'.

(*Page 6, Appendix A, clause A-1.1*) — Add at the end of the sentence:

'(Similar range of thermometers with longer length can also be used so that the overall dimension of the instrument does not exceed 270 + 30 mm, maximum)'.

(*Page 11, clause A-8.1, line 1*) — Substitute 'Psychrometer' for 'thermometer'.

(*Page 11, clause A-9.1, line 1*) — Substitute 'Psychrometer' for 'thermometer'.