



AMENDMENT NO. 1 SEPTEMBER 1982

TO

IS:6905-1973 SPECIFICATION FOR INSTRUMENTS
TABLE. MAYO'S TYPE

Alterations

(Page 1, clause 2.3) - Substitute the following
for the existing matter:

'2.3 Top - Shall be according to 2.2(a), 2.2(b) and
2.2(c) of IS:4033-1968.'

(Page 1, clause 4.2, line 2) - Substitute '1230 mm'
for '1270 mm'.

(Page 2, Fig. 1, dimension for overall height) -
Substitute '830 to 1230' for '830 to 1270'.

(CPDC 14)